

Fax +9/2-4628 82// Email: mail@hermonlabs.com

12.10.2005

To whom it may concern:

Hermon Laboratories has tested Sysmteric's Gravimetric Doser model CD800 according to IEC/EN 61010-1 standard, and found it to be in compliance with the standard's requirements.

Sysmetric has declared that models CD100 and CD400 are reduced size and power version of model CD800 mentioned above

Our conclusion is, (according to Sysmetric's declaration), That models CD100 and CD400 also meets IEC/EN 61010-1 requirements.

Ami Friedman Technical manager

Description of the second